


Measurement wear scar depth on the meniscus



Measurement of the wear scar depth on the meniscus using an Alicona IF G5 optical microscope from Accutrans replicas:

A: Shows a measurement of the meniscus replica and the position of a trace through the wear scar; B: shows the height difference between the actual surface and the reference (unworn surface) created by fitting a sphere to the unworn edges; C: the difference between the worn and unworn surface (ΔZ).