



Name - Calum J Gibb
Room No. - 1.33
Sample - CJG 116

130.81
129.14
127.18
127.11

77.35
77.03
76.71

35.69
31.85
31.48
29.33
29.21
22.70
14.12

Current Data Parameters
NAME CJG 116 RM 1.33
EXPNO 3
PROCNO 1

F2 - Acquisition Parameters
Date_ 20231128
Time 20.15 h
INSTRUM spect
PROBHD Z116098_0395 (
PULPROG zgpg30
TD 32768
SOLVENT CDC13
NS 1024
DS 8
SWH 24038.461 Hz
FIDRES 1.467191 Hz
AQ 0.6815744 sec
RG 201.96
DW 20.800 usec
DE 6.50 usec
TE 298.0 K
D1 1.00000000 sec
D11 0.03000000 sec
TD0 1
SF01 100.6228298 MHz
NUC1 13C
P0 3.33 usec
P1 10.00 usec
PLW1 60.02999878 W
SF02 400.1316005 MHz
NUC2 1H
CPDPRG[2] waltz16
PCPD2 90.00 usec
PLW2 17.06500053 W
PLW12 0.21067999 W
PLW13 0.10597000 W

F2 - Processing parameters
SI 65536
SF 100.6127685 MHz
WDW EM
SSB 0
LB 2.00 Hz
GB 0
PC 1.40

